



Product Service

Attestation of Conformity

No. N8A 114495 0006 Rev. 00

Holder of Attestation: Anhui Wanyi Science and Technology Co., Ltd.

No.8 Wenqu Road, High-tech Industrial Development Zone
230088 Hefei, Anhui
PEOPLE'S REPUBLIC OF CHINA

**Product: Laboratory Equipment
Helium Leak detector**

Model(s): SFJ-231

Parameters:


Rated voltage	230VAC
Rated frequency	50/60Hz
Rated power	1100W
Protection class	I
Degree of protection	IP20

Tested according to: EN 61010-1:2010/A1:2019

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: 4842024026200

Date, 2024-04-23


(Bin Li)

Page 1 of 1

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.